


AG11.2 Refractometer Calibration Fluid Refractive Index Temperature Relationship

Temperature (°C)	0.0	0.1	0.2	0.3	0.4	0.5	0.6	0.7	0.8	0.9
10	1.35076	1.35075	1.35074	1.35073	1.35072	1.35071	1.35070	1.35069	1.35068	1.35067
11	1.35066	1.35065	1.35064	1.35063	1.35062	1.35061	1.35060	1.35059	1.35058	1.35057
12	1.35056	1.35055	1.35054	1.35053	1.35052	1.35051	1.35050	1.35049	1.35048	1.35047
13	1.35046	1.35045	1.35044	1.35043	1.35042	1.35041	1.35040	1.35039	1.35038	1.35037
14	1.35036	1.35035	1.35034	1.35033	1.35032	1.35031	1.35029	1.35028	1.35027	1.35026
15	1.35025	1.35024	1.35023	1.35022	1.35021	1.35020	1.35019	1.35018	1.35016	1.35015
16	1.35014	1.35013	1.35012	1.35011	1.35010	1.35009	1.35008	1.35007	1.35005	1.35004
17	1.35003	1.35002	1.35001	1.35000	1.34999	1.34998	1.34996	1.34995	1.34994	1.34993
18	1.34992	1.34991	1.34989	1.34988	1.34987	1.34986	1.34985	1.34984	1.34982	1.34981
19	1.34980	1.34979	1.34978	1.34977	1.34975	1.34974	1.34973	1.34972	1.34971	1.34969
20	1.34968	1.34967	1.34966	1.34965	1.34963	1.34962	1.34961	1.34960	1.34959	1.34957
21	1.34956	1.34955	1.34954	1.34952	1.34951	1.34950	1.34949	1.34947	1.34946	1.34945
22	1.34944	1.34942	1.34941	1.34940	1.34939	1.34937	1.34936	1.34935	1.34934	1.34932
23	1.34931	1.34930	1.34928	1.34927	1.34926	1.34924	1.34923	1.34922	1.34921	1.34919
24	1.34918	1.34917	1.34915	1.34914	1.34913	1.34911	1.34910	1.34909	1.34907	1.34906
25	1.34905	1.34903	1.34902	1.34901	1.34899	1.34898	1.34897	1.34895	1.34894	1.34893
26	1.34891	1.34890	1.34888	1.34887	1.34886	1.34884	1.34883	1.34881	1.34880	1.34879
27	1.34877	1.34876	1.34874	1.34873	1.34872	1.34870	1.34869	1.34867	1.34866	1.34865
28	1.34863	1.34862	1.34860	1.34859	1.34857	1.34856	1.34855	1.34853	1.34852	1.34850
29	1.34849	1.34847	1.34846	1.34844	1.34843	1.34841	1.34840	1.34838	1.34837	1.34835
30	1.34834	1.34832	1.34831	1.34829	1.34828	1.34826	1.34825	1.34823	1.34822	1.34820
31	1.34819	1.34817	1.34816	1.34814	1.34813	1.34811	1.34810	1.34808	1.34807	1.34805
32	1.34804	1.34802	1.34800	1.34799	1.34797	1.34796	1.34794	1.34793	1.34791	1.34789
33	1.34788	1.34786	1.34785	1.34783	1.34781	1.34780	1.34778	1.34777	1.34775	1.34773
34	1.34772	1.34770	1.34769	1.34767	1.34765	1.34764	1.34762	1.34760	1.34759	1.34757
35	1.34755	1.34754	1.34752	1.34750	1.34749	1.34747	1.34745	1.34744	1.34742	1.34740
36	1.34739	1.34737	1.34735	1.34734	1.34732	1.34730	1.34729	1.34727	1.34725	1.34723
37	1.34722	1.34720	1.34718	1.34717	1.34715	1.34713	1.34711	1.34710	1.34708	1.34706
38	1.34704	1.34703	1.34701	1.34699	1.34697	1.34696	1.34694	1.34692	1.34690	1.34689
39	1.34687	1.34685	1.34683	1.34681	1.34680	1.34678	1.34676	1.34674	1.34672	1.34670
40	1.34669									

AG11.2 Refractometer Calibration Fluid - R.I. Temperature Relationship



Code: 90-307 Issue: 1



Bellingham + Stanley Ltd.

Longfield Road, Tunbridge Wells

KENT. TN2 3EY

UK

Telephone: +44 (0) 1892 500400

Fax: +44 (0) 1892 543115

Bellingham + Stanley Inc.

1000 Hurricane Shoals Road

Bldg D, Suite 300

LAWRENCEVILLE, USA, GA 30043


Telephone: 770 822 6898

Fax: 770 822 9165

AG11.2 Refractometer Calibration Fluid °Brix Temperature Relationship

Temperature (°C)	0.0	0.1	0.2	0.3	0.4	0.5	0.6	0.7	0.8	0.9
10	11.89	11.88	11.88	11.87	11.86	11.86	11.85	11.84	11.84	11.83
11	11.83	11.82	11.81	11.81	11.80	11.79	11.79	11.78	11.77	11.77
12	11.76	11.76	11.75	11.74	11.74	11.73	11.72	11.72	11.71	11.70
13	11.70	11.69	11.68	11.68	11.67	11.66	11.66	11.65	11.64	11.64
14	11.63	11.62	11.62	11.61	11.60	11.60	11.59	11.58	11.58	11.57
15	11.56	11.56	11.55	11.54	11.53	11.53	11.52	11.51	11.51	11.50
16	11.49	11.49	11.48	11.47	11.46	11.46	11.45	11.44	11.44	11.43
17	11.42	11.41	11.41	11.40	11.39	11.38	11.38	11.37	11.36	11.36
18	11.35	11.34	11.33	11.33	11.32	11.31	11.30	11.30	11.29	11.28
19	11.27	11.27	11.26	11.25	11.24	11.24	11.23	11.22	11.21	11.20
20	11.20	11.19	11.18	11.17	11.17	11.16	11.15	11.14	11.13	11.13
21	11.12	11.11	11.10	11.09	11.09	11.08	11.07	11.06	11.05	11.05
22	11.04	11.03	11.02	11.01	11.01	11.00	10.99	10.98	10.97	10.96
23	10.96	10.95	10.94	10.93	10.92	10.92	10.91	10.90	10.89	10.88
24	10.87	10.86	10.86	10.85	10.84	10.83	10.82	10.81	10.80	10.80
25	10.79	10.78	10.77	10.76	10.75	10.74	10.74	10.73	10.72	10.71
26	10.70	10.69	10.68	10.67	10.66	10.66	10.65	10.64	10.63	10.62
27	10.61	10.60	10.59	10.58	10.57	10.57	10.56	10.55	10.54	10.53
28	10.52	10.51	10.50	10.49	10.48	10.47	10.46	10.45	10.44	10.44
29	10.43	10.42	10.41	10.40	10.39	10.38	10.37	10.36	10.35	10.34
30	10.33	10.32	10.31	10.30	10.29	10.28	10.27	10.26	10.25	10.24
31	10.23	10.22	10.21	10.20	10.19	10.18	10.17	10.16	10.15	10.14
32	10.13	10.12	10.11	10.10	10.09	10.08	10.07	10.06	10.05	10.04
33	10.03	10.02	10.01	10.00	9.99	9.98	9.97	9.96	9.95	9.94
34	9.93	9.92	9.91	9.90	9.89	9.87	9.86	9.85	9.84	9.83
35	9.82	9.81	9.80	9.79	9.78	9.77	9.76	9.75	9.73	9.72
36	9.71	9.70	9.69	9.68	9.67	9.66	9.65	9.64	9.62	9.61
37	9.60	9.59	9.58	9.57	9.56	9.55	9.53	9.52	9.51	9.50
38	9.49	9.48	9.47	9.45	9.44	9.43	9.42	9.41	9.40	9.39
39	9.37	9.36	9.35	9.34	9.33	9.32	9.30	9.29	9.28	9.27
40	9.26									

AG11.2 Refractometer Calibration Fluid - °Brix Temperature Relationship



Code: 90-307 Issue: 1



Bellingham + Stanley Ltd.

Longfield Road, Tunbridge Wells
 KENT. TN2 3EY
 UK
 Telephone: +44 (0) 1892 500400
 Fax: +44 (0) 1892 543115

Bellingham + Stanley Inc.

1000 Hurricane Shoals Road
 Bldg D, Suite 300
 LAWRENCEVILLE, USA, GA 30043
 Telephone: 770 822 6898
 Fax: 770 822 9165