

REFRACTIVE INDEX & °BRIX TEMPERATURE RELATIONSHIP

Calibration Oil Component Code: 90-225 Type: BSLP Batch: BSLP23

TEMPERATURE	REFRACTIVE INDEX	°BRIX	TEMPERATURE	REFRACTIVE INDEX	°BRIX
20.0°C	1.46990	71.81	50.0°C	1.45875	67.22
25.0°C	1.46804	71.06	55.0°C	1.45689	66.44
30.0°C	1.46618	70.30	60.0°C	1.45503	65.66
35.0°C	1.46433	69.54	65.0°C	1.45317	64.87
40.0°C	1.46247	68.77	70.0°C	1.45131	64.08
45.0°C	1.46061	68.00	75.0°C	1.44945	63.28
R.I. = (1.477321E+00) + (-3.709934E-04 x Temperature) + (-8.313294E-09 x Temperature ²)					
°BRIX = (7.478665E+01) + (-1.469925E-01 x Temperature) + (-8.561439E-05 x Temperature ²)					

Refractive index values detailed in the table were calculated from a 2nd order polynomial fit of empirical data, as shown above, which is valid over the range 20 to 75°C. Values used to determine this relationship were produced from a statistical evaluation of averaged refractive index data.

Equivalent °Brix values are derived from ICUMSA SPS-3 (2000) Equation 2. When using these values to calibrate at the refractometer's operating temperature, sucrose temperature compensation should not be applied to °Brix readings.

The refractive index of the sample was measured on a high accuracy refractometer at 589.3nm in a temperature controlled room. Refractometer calibration was set prior to testing with high purity water. Refractive index values for water were obtained from "Revised Formulation for the Refractive Index of Water and Steam as a Function of Wavelength, Temperature and Density", adopted by the International Association for the Properties of Water and Steam (IAPWS) and available as part of NIST Standard Reference Database 10¹. Refractive indices calculated from the formulation are absolute refractive indices; conversion to refractive index against air requires division by the respective absolute refractive index of air (NIST Engineering Metrology Toolbox²).

Maximum residual of polynomial data fit: ±0.00005 R.I.

Date of analysis: 22nd November 2007

Authorised signatory



A. Darkins
Laboratory Manager

Calibration oils are guaranteed until the expiry date, provided they are stored as directed. However, B+S can only consider queries regarding a batch of samples if unopened bottles are returned before the expiry date. In no event will Bellingham + Stanley Ltd. be liable for direct, indirect, incidental or consequential damages resulting from the use of this product.

Material Safety Data Sheet for this product is available from the B+S website or on request.

Equivalent °Brix values and refractive indices at temperatures other than 20.0°C are not covered by the current UKAS Scope of Accreditation of Bellingham + Stanley Ltd. °Brix values greater than 85.00 are extrapolated from the relationship given in ICUMSA SPS-3 (2000) Equation 2.

¹NIST Standard Reference Database 10, Version 2.21, A.H.Harvey, A.P.Peskin, S.A.Klein

²NIST Engineering Metrology Toolbox: Refractive Index of Air Calculator (<http://emtoolbox.nist.gov/Wavelength/Abstract.asp>)

www.bellinghamandstanley.com

Part Code: 90-370 Issue: 03



Bellingham + Stanley Ltd.

Longfield Road, Tunbridge Wells

KENT. TN2 3EY

UK

Telephone: +44 (0) 1892 500400

Fax: +44 (0) 1892 543115